

# ANALYSENZERTIFIKAT

## CERTIFICATE OF ANALYSIS

<b>SERVA Silver Stain Kit</b> for SDS PAGE	<b>Kat.Nr./Cat.No. :</b> 35076 <b>Lot/Contr.No. :</b> P200371
---	--

Parameter parameter	Spezifikation specification	Ergebnis result
<b>Anwendung</b> application	SDS PAGE	entspricht corresponds
<b>Inhalt</b> content	500 ml Fixing solution 25 x 30mg Natriumthiosulfate 5 H <sub>2</sub> O 250 ml Silver nitrate solution 1.5 ml Formaldehyde 500 ml Sodium carbonate solution 30 g Glycine	
<b>Mindesthaltbarkeit</b> minimum shelf life		26.06.2021
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Bei Lagerung unter +10 °C kann die Natriumcarbonat Lösung auskristallisieren.  
Die Kristalle lösen sich beim Erwärmen des Produktes wieder.**

Storage below +10 °C may cause crystallization of Sodium carbonate solution.  
Crystals can be solved by warming up the product.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 21.10.2020

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.