

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVA Silver Stain Kit
for SDS PAGE

Kat.Nr./Cat.No. : 35076
Lot/Contr.No. : P240062

Parameter parameter	Spezifikation specification	Ergebnis result
Anwendung application	SDS PAGE	entspricht corresponds
Inhalt content	500 ml Fixing solution 25 x 30mg Sodiumthiosulfate 5 H ₂ O 250 ml Silver nitrate solution 1.5 ml Formaldehyde 500 ml Sodium carbonate solution 30 g Glycine	
Mindesthaltbarkeit minimum shelf life		31.03.2025
Lagerung (°C) storage		+15 bis/ to +30

**Bei Lagerung unter +10 °C kann die Natriumcarbonat Lösung auskristallisieren.
Die Kristalle lösen sich beim Erwärmen des Produktes wieder.**

Storage below +10 °C may cause crystallization of Sodium carbonate solution.
Crystals can be solved by warming up the product.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.02.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.