

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVA ProteinStain Fluo-R	Kat.Nr./Cat.No. :	35090
SERVA ProteinStain Fluo-R	Lot/Contr.No. :	221239

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		$C_{72}H_{42}N_6O_{18}S_6 \cdot Na_4Ru \cdot nH_2O$	
Molekulargewicht molecular weight	wasserfrei anhydrous	1664.55	
Aussehen appearance		orange-rotes Pulver orange-red powder	entspricht corresponds
Identität identity	1H -NMR	entspricht corresponds	entspricht corresponds
Reinheit purity	DC TLC	ein Spot single spot	entspricht corresponds
Anwendung application	Proteindetektion in SDS-PAGE protein detection in SDS-PAGE	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			16.12.2027
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 19.12.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.