

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Siliconöl DC 550; 115 cst pract.	Kat.Nr./Cat.No. : 35145
Silicone DC 550 fluid; 115 cst pract.	Lot/Contr.No. : 230821

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, viskose Flüssigkeit clear, colorless liquid	entspricht corresponds
Dichte (g/ml) specific gravity		ca. 1.06	entspricht corresponds
Brechungsindex refractive index	25°C	1.4900 - 1.500	1.4997
Viskosität (cst) viscosity		100.0 – 150.0	140.6
Mindesthaltbarkeit minimum shelf life			18.02.2028
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.07.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.