

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                              |                                |
|----------------------------------------------|--------------------------------|
| <b>Siliconöl DC 710; 500 cst pract.</b>      | <b>Kat.Nr./Cat.No. : 35149</b> |
| <b>Silicone DC 710 fluid; 500 cst pract.</b> | <b>Lot/Contr.No. : 210054</b>  |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method | <b>Spezifikation</b><br>specification                 | <b>Ergebnis</b><br>result |
|-------------------------------------------------|--------------------------|-------------------------------------------------------|---------------------------|
| <b>Aussehen</b><br>appearance                   |                          | klare, viskose Flüssigkeit<br>clear, colorless liquid | entspricht<br>corresponds |
| <b>Dichte (g/ml)</b><br>specific gravity        | 25 °C                    | 1.10 – 1.11                                           | entspricht<br>corresponds |
| <b>Brechungsindex</b><br>refractive index       | 25 °C                    | 1.5300 – 1.5380                                       | 1.5348                    |
| <b>Viskosität (cst)</b><br>viscosity            | 25 °C                    | 450.0 – 550.0                                         | 523.9                     |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                          |                                                       | 02.02.2025                |
| <b>Lagerung (°C)</b><br>storage                 |                          |                                                       | +15 bis / to +30          |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 21.01.2021**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.