

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Spermidin-3HCl reinst	Kat.Nr./Cat.No. :	35285
Spermidine-3HCl research grade	Lot/Contr.No. :	121156

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₇ H ₁₉ N ₃ ·3HCl	
Molekulargewicht molecular weight		254.6	
Aussehen appearance		weiße Kristalle white crystals	entspricht corresponds
Gehalt (%) assay	Titration	min. 99.0	99.9
Schmelzpunkt (°C) melting point			257
Identität identity	IR	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			12/ 2016
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.10.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.