

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Spermidin-3HCl reinst	Kat.Nr./Cat.No. :	35285
Spermidine-3HCl research grade	Lot/Contr.No. :	220265

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₇ H ₁₉ N ₃ ·3HCl	
Molekulargewicht molecular weight		254.6	
Aussehen appearance		weiße Kristalle white crystals	entspricht corresponds
Gehalt (%) assay	Titration	min. 99.0	100.5
Schmelzpunkt (°C) melting point			257
Identität identity	IR	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			10.03.2026
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.03.2022

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.