

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Synperonic® F68 pract.	Kat.Nr./Cat.No. : 35724
	Lot/Contr.No. : 130060

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Molekulargewicht molecular weight		ca. 8300	entspricht corresponds
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
pH	EN 1262 Lösung/ solution B	5.0 – 8.0	6.7
pH	5 % in H ₂ O	5.0 – 8.0	6.0
Identität identity	IR	entspricht corresponds	entspricht corresponds
Trübungspunkt (°C) cloud point	EN 1890 Verfahren/ method B (in 5 % NaCl Lösung/ solution)	86.0 – 90.0	86.7
Iod-Farbzahl iodine colour	EN 1557, 70 °C	max. 2.0	1.6
Ursprung origin		nicht tierisch non animal	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			01/ 2015
Lagerung (°C) storage			+15 bis/ to +30

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.01.2013

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.