

**ANALYSENZERTIFIKAT**  
 CERTIFICATE OF ANALYSIS

|                               |                          |               |
|-------------------------------|--------------------------|---------------|
| <b>Synperonic® F68 pract.</b> | <b>Kat.Nr./Cat.No. :</b> | <b>35724</b>  |
|                               | <b>Lot/Contr.No. :</b>   | <b>200342</b> |

| Parameter<br>parameter                          | Methode<br>method   | Spezifikation<br>specification | Ergebnis<br>result        |
|---|---|--------------------------------|---------------------------|
| <b>Molekulargewicht</b><br>molecular weight     |   | ca. 8300                       | entspricht<br>corresponds |
| <b>Aussehen</b><br>appearance                   |   | weißes Pulver<br>white powder  | entspricht<br>corresponds |
| <b>pH</b>                                       | EN 1262<br>Lösung/ solution B                                 | 5.0 – 8.0                      | 6.0                       |
| <b>Identität</b><br>identity                    | IR  | entspricht<br>corresponds      | entspricht<br>corresponds |
| <b>Trübungspunkt (°C)</b><br>cloud point        | EN 1890 Verfahren/ method B<br>(in 5 % NaCl Lösung/ solution) | 86.0 – 90.0                    | 86.0                      |
| <b>Iod-Farbzahl</b><br>iodine colour            | EN 1557, 70 °C  | max. 2.0                       | 1                         |
| <b>Ursprung</b><br>origin                       |   | nicht tierisch<br>non animal   | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |   |                                | 01.04.2022                |
| <b>Lagerung (°C)</b><br>storage                 |   |                                | +15 bis/ to +30           |

|  |
|--|
| <p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b><br/> <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b><br/>         We do not guarantee that the product can be used for a special application.<br/>         This document does not release you from performing the standard control upon receipt of incoming goods.</p> |
|--|

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date:** 06.04.2020

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.