

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Trichloressigsäure p.A.	Kat.Nr./Cat.No. :	36910
Trichloracetic acid analytical grade	Lot/Contr.No. :	250982

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₂ HCl ₃ O ₂	
Molekulargewicht molecular weight		163.40	
Aussehen appearance		weiße Kristalle white crystals	entspricht corresponds
Gehalt (%) assay	Titration	min. 99.0	99.9
Trocknungsverlust (%) loss on drying		max. 0.5	0.1
Schmelzpunkt (°C) melting point		min. 57.5	57.9
Sulfatasche (%) sulfated ash		max. 0.03	< 0.01
Schwermetalle (ppm) heavy metals		max. 20	< 20
Mindesthaltbarkeit minimum shelf life			19.05.2027
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.10.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.