

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>TMB, gebrauchsfertiges ELISA-Substrat</b>	<b>Kat.Nr./Cat.No. : 37068</b>
TMB, ready-to-use ELISA substrate	<b>Lot/Contr.No. : 161042</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare farblose bis leicht blaue Lösung clear colourless to slightly blue solution	entspricht corresponds
<b>pH</b>		3.75 – 3.85	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			30.09.2018
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 07.11.2016**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.