

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

SERVAColor TMB Blot-Lösung	Kat.Nr./Cat.No. : 37071
SERVAColor TMB Blot Solution	Lot/Contr.No. : 170176

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		farblose bis leicht gelbe Flüssigkeit colourless to pale yellow liquid	entspricht corresponds
pH		4.9 – 5.1	5.0
Mindesthaltbarkeit minimum shelf life			30.09.2021
Lagerung (°C) storage			+2 bis/ to +8

SERVA Electrophoresis GmbH

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 23.06.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.