

CERTIFICATE OF ANALYSIS

Protease-Inhibitor-Mix M (Cat.No. :	39102
	Lot No.:	P130001

Cat. number	Components	Target Protease
12745	AEBSF-HCl	Serine Proteases
13718	Aprotinin from bovine lung	Serine Proteases
14980	Bestatin	Aminopeptidase B and Leucine Aminopeptidase
21100	E-64	Cysteine Proteases
51867	Leupeptin	Cysteine Proteases and Trypsin-like Serine Proteases
52682	Pepstatin A	Aspartic Proteases
Storage (°C)	R.	-15 to -25
Minimum shelf life		01/2016

ReconstitutionDissolve the content of the vial (15.8mg) in 1ml DMSO to obtain a 100x
stock solution (suitable for the treatment of 100ml tissue extract).
After reconstitution, aliquot and freeze at -20°C (stable for at least 4
weeks).

We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH Quality Control

Printing date: 03.01.2013

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

This report has been computer generated and does not contain a signature.

SERVA Electrophoresis GmbH • D-69115 Heidelberg • Carl-Benz-Str. 7 • Tel.: +49 6221 138 40-0 • Fax +49 6221 138 40-10 Geschäftsführer: Dr. Barbara Müller, Dr. Manfred Kurfürst • HRB 336136 AG Mannheim • E-Mail: info.serva.de • http://www.serva.de COMMERZBANK BLZ: 672 400 39 Kto: 190 38 30 00 • IBAN: DE90 6724 0039 0190 3830 00 • BIC: COBADEFFXXX • Ust.ID-Nr.: DE 812 517 285