

Certificate of Analysis

Protein Test Mixture 6 for SDS PAGE

Cat.No. : 39207

Lot No. : P210115

Protein Components	Molecular weight
Phosphorylase B	97 400
Albumin bovine (BSA)	67 000
Albumin egg	45 000
Carbonic anhydrase	29 000
Trypsin inhibitor (soybean)	21 000
Cytochrome C	12 300
Trypsin inhibitor (bovine lung)	6 500
Minimum shelf life	13.04.2023
Storage (°C)	+2 to +8

Reconstitution

Dissolve to a concentration of 1 mg/ ml in SDS-sample buffer.
A volume of 5 µl is sufficient for SDS-PAGE (in a standard form of 80 x 60 x 0.75 mm).

SDS-sample buffer:

125 mM Tris-HCl pH 6.8; 2% SDS; 15% glycerol;
10 mM DTT; 0.025% Bromphenol blue; 0.025% Orange G

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 14.04.2021

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.