

CERTIFICATE OF ANALYSIS**Protein Test Mixture 4****Cat.No. : 39208****Lot No.: P190392****Protein Components****Molecular weight****Carbonic anhydrase**

29 000

Albumin egg

45 000

Albumin bovine

67 000

Phosphorylase B

97 400

Minimum shelf life

22.11.2021

Storage (°C)

+2 to +8

Reconstitution

Dissolve to a concentration of 1mg/ml in SDS-sample buffer.
A volume of 5µl is sufficient for SDS-PAGE (in a standard
form of 80 x 60 x 0.75mm).

SDS-sample buffer:

125 mM Tris-HCl pH 6.8; 2% SDS; 15% glycerol;
10 mM DTT; 0.025% Bromphenol blue; 0.025% Orange G

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Quality Control**Printing date: 22.11.2019**

Christian Monsler

Marina Konter

This report has been computer generated and does not contain a signature.