

CERTIFICATE OF ANALYSIS

Protein Test Mixture 4

Cat.No. : 39208

Lot No.: P220062

Protein Components

Molecular weight

Carbonic anhydrase

29 000

Albumin egg

45 000

Albumin bovine

67 000

Phosphorylase B

97 400

Minimum shelf life

25.02.2024

Storage (°C)

+2 to +8

Reconstitution

Dissolve to a concentration of 1mg/ml in SDS-sample buffer.
A volume of 5µl is sufficient for SDS-PAGE (in a standard form of 80 x 60 x 0.75mm).

SDS-sample buffer:

125 mM Tris-HCl pH 6.8; 2% SDS; 15% glycerol;
10 mM DTT; 0.025% Bromphenol blue; 0.025% Orange G

We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Quality Control

Printing date: 02.03.2022

Christian Monsler

Daniela Lux-Helmstetter

This report has been computer generated and does not contain a signature.