

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

BCA Protein Assay Macro Kit	Kat.Nr./Cat.No. :	39228
BCA Protein Assay Macro Kit	Lot/Contr.No. :	P230182

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Anwendung application	BCA Protein Assay	entspricht corresponds	entspricht corresponds
Komponenten components		Reagent A: BCA-Lösung Reagent B: Kupfersulfat-Lösung Proteinstandard (Rinderserumalbumin) Reagent A: BCA solution Reagent B: copper sulphate solution Protein standard (bovine serum albumin)	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			24.10.2024
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 05.05.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.