

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>RIPA-Puffer</b>	<b>Kat.Nr./Cat.No. : 39244</b>
RIPA Buffer	<b>Lot/Contr.No. : P210223</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	klare, farblose Lösung clear, colourless solution	entspricht corresponds
<b>Zusammensetzung</b> composition		
Tris buffer pH 7.5	25 mM	entspricht/ corresponds
NaCl	150 mM	entspricht/ corresponds
Tergitol™ 15-S-9	1%	entspricht/ corresponds
Deoxycholic acid-Na-salt	1%	entspricht/ corresponds
SDS	0.1%	entspricht/ corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		28.06.2023
<b>Lagerung (°C)</b> storage		+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 30.06.2021**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.