

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

RIPA-Puffer	Kat.Nr./Cat.No. : 39244
RIPA Buffer	Lot/Contr.No. : P240043

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colourless solution	entspricht corresponds
Zusammensetzung composition		
Tris buffer pH 7.5	25 mM	entspricht/ corresponds
NaCl	150 mM	entspricht/ corresponds
Tergitol™ 15-S-9	1%	entspricht/ corresponds
Deoxycholic acid-Na-salt	1%	entspricht/ corresponds
SDS	0.1%	entspricht/ corresponds
Mindesthaltbarkeit minimum shelf life		08.02.2026
Lagerung (°C) storage		+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.02.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.