

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Chloroform	Kat.Nr./Cat.No. :	39553
molecular biology grade	Lot/ Contr.No. :	180947

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		CHCl ₃	
Molekulargewicht molecular weight		119.39	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Gehalt (%) assay	GC	99.9 – 100.0	99.99
Dichte (g/ml) specific density		1.478 – 1.482	entspricht corresponds
Brechungsindex refractive index	n _D ²⁰		1.4449
Wasser (%) water	KF	max. 0.05	0.007
Säure (%) acidity		max. 0.001	0.0003
Abdampfrückstand (w/w %) residue on evaporation		max. 0.0005	0.0004
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			31.07.2021
Lagerung (°C) storage			+15 bis/ to +30

Das Produkt ist stabilisiert mit ca. 1 % Ethanol.
The product is stabilized with approx. 1 % ethanol.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.09.2018

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.