

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

---

<b>Chloroform:Isoamyl alcohol 24:1</b> molecular biology grade	<b>Kat.Nr./Cat.No. : 39554</b> <b>Lot/ Contr.No. : P200052</b>
-------------------------------------------------------------------	-------------------------------------------------------------------

---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Reinheit des Chloroform(%)</b> purity of Chloroform		min. 99.9	> 99.9
<b>Reinheit des Isoamyl alcohol (%)</b> purity of Isoamyl alcohol		min. 99.0	entspricht corresponds
<b>Brechungsindex</b> refractive index	$n_D^{20\text{ °C}}$		1.4432
<b>DNase/ RNase Aktivität</b> DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			07.02.2023
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

---

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 10.02.2020**

Christian Monsler

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.