

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Ethanol unvergällt absolut	Kat.Nr./Cat.No. :	39556
Ethanol undenaturated absolute molecular biology grade	Lot/Contr.No. :	131265

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₂ H ₆ O	
Molekulargewicht molecular weight		46.07	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Gehalt (%) assay	GC	min. 99.7	100.0
Wasser (%) water	KF	max. 0.2	0.04
Brechungsindex refractive index	n _D / 20 °C	1.360 – 1.362	1.3613
Identität identity	IR	entspricht corresponds	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			01/ 2018
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 18.11.2013

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.