

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Isopropanol	Kat.Nr./Cat.No. :	39559
molecular biology grade	Lot/ Contr.No. :	160090

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₃ H ₈ O	
Molekulargewicht molecular weight		60.09	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Gehalt (%) assay	GC	min. 99.7	entspricht corresponds
Dichte (g/ml) specific density	20 °C	0.785 – 0.786	entspricht corresponds
Brechungsindex refractive index	20 °C	1.376 – 1.378	1.3771
Wasser (%) water		max. 0.1	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			20.01.2020
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.01.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.