

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

2-Mercaptoethanol	Kat.Nr./Cat.No. :	39563
molecular biology grade	Lot/Contr.No. :	160871

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₂ H ₆ OS	
Molekulargewicht molecular weight		78.13	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless solution	entspricht corresponds
Gehalt (%) assay	iod. Titration iod. titration	min. 99.0	99.7
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			31.12.2018
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 06.11.2017

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.