

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

20% SDS Lösung	Kat.Nr./Cat.No. : 39575
20% SDS solution, molecular biology grade	Lot/Contr.No. : P220040

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose bis leicht gelbliche Lösung clear, colorless to slightly yellowish solution	entspricht corresponds
Gehalt SDS (g/l) assay SDS	200.0	entspricht corresponds
Farbe (Hazen) color	max. 50	5
A 1 cm/ 260 nm	max. 0.4	0.027
A 1 cm/ 280 nm	max. 0.15	0.019
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		11.02.2023
Lagerung (°C) storage		+15 bis/ to +30

Bei Lagerung unter +18°C kann das SDS auskristallisieren.
Die Kristalle lösen sich beim Erwärmen des Puffers wieder.
Storage below +18°C may cause crystallization of SDS.
Crystals can be solved by warming up the buffer solution.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control
Druckdatum/ printing date: 22.02.2022

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.