

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Guanidin-thiocyanat	Kat.Nr./Cat.No. :	39577
Guanidine-thiocyanate molecular biology grade	Lot/Contr.No. :	221016

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		CH ₅ N ₃ ·HSCN	
Molekulargewicht molecular weight		118.16	
Aussehen appearance		farblose bis weiße Kristalle oder kristallines Pulver colourless to white crystals or crystalline powder	entspricht corresponds
Gehalt (%) assay		min. 98.5	99.8
Identität identity	IR	entspricht corresponds	entspricht corresponds
Löslichkeit solubility	20 % in H ₂ O	klar und farblos clear and colourless	entspricht corresponds
Schmelzpunkt (°C) melting point		115 – 122	120
Wasser/ water (%)	Karl-Fischer-Titration	max. 0.5	0.2
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			18.10.2026
Lagerung (°C) storage			+15 bis / to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.10.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.