

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | |
|---|--|
| Glycogen from oyster, solution 20 mg/ml molecular biology grade | Kat.Nr./Cat.No. : 39766 Lot/Contr.No. : P190044 |
|---|--|

| Parameter parameter | Spezifikation specification | Ergebnis result |
|--|---------------------------------------|---------------------------|
| Aussehen appearance | milchige Lösung cloudy solution | entspricht corresponds |
| Konzentration concentration | 20 mg/ ml in redistilled water | entspricht corresponds |
| DNase/ RNase Aktivität DNase/ RNase activity | nicht nachweisbar none detected | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 23.01.2021 |
| Lagerung (°C) storage | | -15 bis/ to -25 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control**Druckdatum/ printing date: 12.02.2019**

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.