

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Tris-Puffer pH 8.0, 1M Lösung	Kat.Nr./Cat.No. : 39792
Tris buffer pH 8.0, solution 1M molecular biology grade	Lot/Contr.No. : P200225

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare farblose Lösung clear colorless solution	entspricht corresponds
Zusammensetzung composition			
TRIS (#37186)		121.14 g/l	entspricht corresponds
pH	20°C eingestellt mit HCl adjusted with HCl	7.9 - 8.1	8.1
DNase/ RNase Aktivität DNase/ RNase activity		nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			30.06.2022
Lagerung (°C) storage			+15 bis / to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 07.07.2020

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.