

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

TE Puffer (100x) pH 8.0 molecular biology grade	Kat.Nr./Cat.No. : 39799
TE buffer (100x) pH 8.0	Lot/Contr.No. : P200271

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	7.8 - 8.2	8.2
Zusammensetzung composition		
TRIS (#37180)	1.0 M	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O (# 11280)	0.1 M	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		03.08.2022
Lagerung (°C) storage		+15 bis / to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 10.08.2020

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.