

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Annexin V-FITC Apoptose-Detektions-Kit	Kat.Nr./Cat.No. :	39900
Annexin V-FITC Apoptosis Detection Kit	Lot/Contr.No. :	200280

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen Annexin V-FITC Lösung appearance Annexin V-FITC solution	klare, farblose Lösung clear, colourless solution	entspricht corresponds
Aussehen Bindungspuffer (10x) appearance binding buffer (10x)	klare, farblose Lösung clear, colourless solution	entspricht corresponds
Aussehen Propidiumiodid Lösung appearance propidium iodide solution	klare, orangene bis rote Lösung clear, orange to red solution	entspricht corresponds
Anwendung application	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		31.08.2021
Lagerung (°C) storage		+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 19.03.2020

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.