

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>XTT Zell-Proliferations-Assay</b>	<b>Kat.Nr./Cat.No. :</b>	<b>39904</b>
XTT Cell Proliferation Assay	<b>Lot/Contr.No. :</b>	<b>241035</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen XTT-Reagenz</b> appearance XTT reagent	weiße bis gelbliche Lösung white to yellowish solution	entspricht corresponds
<b>Aussehen Aktivierungs-Reagenz</b> appearance activation reagent	gelbe Lösung yellow solution	entspricht corresponds
<b>Anwendung</b> application	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		30.11.2026
<b>Lagerung (°C)</b> storage		-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 25.11.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.