

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Resazurin Zell-Viabilitäts-Assay</b>	<b>Kat.Nr./Cat.No. :</b>	<b>39905</b>
Resazurin Cell Viability Assay	<b>Lot/Contr.No. :</b>	<b>200284</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	dunkelblaue Lösung dark-blue solution	entspricht corresponds
<b>Anwendung</b> application	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		30.04.2022
<b>Lagerung (°C)</b> storage		-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control****Druckdatum/ printing date:** 19.03.2020

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.