

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DOWEX[®] 50 WX2 (100 – 200 mesh) pract.	Kat.Nr./Cat.No. : 41520
	Lot/Contr.No. : 160880

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		graubeiges Korn grayish beige colored particles	entspricht corresponds
Gegenion counter ion		H ⁺	entspricht corresponds
Trocknungsverlust (%) loss on drying	3 h, 105 °C	74.0 – 82.0	80.8
Kapazität (eq/l) capacity		min. 0.6	0.7
Mindesthaltbarkeit minimum shelf life			14.09.2021
Lagerung (°C) storage			+15 bis/ to +30

Wir empfehlen, das Harz vor Gebrauch gründlich mit Wasser zu waschen.
Please wash resin thoroughly with water before use.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 19.09.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.