

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Serdolit® Chelite® P p.A.	Kat.Nr./Cat.No. : 41706
analytical grade	Lot/Contr.No. : P200346

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		hell elfenbeinfarbenes Korn light ivory colored particles	entspricht corresponds
Korngröße (mm) particle size		0.3 – 0.8	entspricht corresponds
Gegenion counter ion		Na ⁺	entspricht corresponds
Trocknungsverlust (%) loss on drying	3 h, 105 °C	60.0 – 70.0	63.2
Kapazität (eq/l) capacity		min. 1.1	1.69
Mindesthaltbarkeit minimum shelf life			25.09.2025
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 01.10.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.