

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

| | |
|--|--------------------------------|
| Laemmlli Probenpuffer (2x), für SDS PAGE | Kat.Nr./Cat.No. : 42526 |
| Laemmlli sample buffer (2x), for SDS PAGE | Lot/Contr.No. : P210349 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---------------------------------------|---------------------------|
| Aussehen appearance | blaue Lösung blue solution | entspricht corresponds |
| pH | 6.6 – 7.0 | 7.0 |
| Zusammensetzung composition | | |
| TRIS/HCl pH 6.8 (# 37180) | 126 mM | entspricht corresponds |
| Glycerin / glycerol (# 23176) | 20 % | entspricht corresponds |
| SDS (# 20763) | 4 % | entspricht corresponds |
| Bromphenolblau (# 15375) bromphenol blue | 0.02 % | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 08.10.2023 |
| Lagerung (°C) storage | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 21.10.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.