

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Laemmli Probenpuffer (2x), für SDS PAGE	Kat.Nr./Cat.No. : 42526
Laemmli sample buffer (2x), for SDS PAGE	Lot/Contr.No. : P240027

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	blaue Lösung blue solution	entspricht corresponds
pH	6.6 – 7.0	7.0
Zusammensetzung composition		
TRIS/HCl pH 6.8 (# 37180)	126 mM	entspricht corresponds
Glycerin / glycerol (# 23176)	20 %	entspricht corresponds
SDS (# 20763)	4 %	entspricht corresponds
Bromphenolblau (# 15375) bromphenol blue	0.02 %	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		23.01.2026
Lagerung (°C) storage		+2 bis/ to +8

<p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control
Druckdatum/ printing date: 05.02.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.