

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Probenpuffer für Blue Native (2x)</b> Sample Buffer for Blue Native (2x)	<b>Kat.Nr./Cat.No. : 42533</b> <b>Lot/Contr.No. : P090191</b>
--	--

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>pH</b>	6.9 – 7.1	7.0
<b>Zusammensetzung</b> composition		
ε-Aminocapronsäure / ε-Aminocaproic acid (#12548)	1 M	entspricht corresponds
Bis(2-hydroxyethylamino]-2-(hydroxy- methyl)- 1,3-propanediol / BISTRIS (#15107)	100 mM	entspricht corresponds
Natriumchlorid / Sodium chloride (#30183)	100 mM	entspricht corresponds
Glycerin / Glycerol (#23176)	20 %	entspricht corresponds
Serva Blau G250/ Serva Blue G250 (#35050)	0.2 %	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		01/ 2013
<b>Lagerung (°C)</b> storage		-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 20.12.2011**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.