

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Probenpuffer für Blue Native (2x)	Kat.Nr./Cat.No. : 42533
Sample Buffer for Blue Native (2x)	Lot/Contr.No. : P140410

Parameter parameter	Spezifikation specification	Ergebnis result
pH	6.9 – 7.1	7.0
Zusammensetzung composition		
ε-Aminocapronsäure/ ε-Aminocaproic acid (#12548)	1 M	entspricht corresponds
Bis(2-hydroxyethyl)amino]-2-(hydroxy- methyl)- 1,3-propanediol/ BISTRIS (#15107)	100 mM	entspricht corresponds
Natriumchlorid/ Sodium chloride (#30183)	100 mM	entspricht corresponds
Glycerin/ Glycerol (#23176)	20 %	entspricht corresponds
Serva Blau G250/ Serva Blue G250 (#35050)	0.2 %	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		03/ 2017
Lagerung (°C) storage		-15 bis/ to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 10.02.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.