

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>Probenpuffer für Blue Native (2x)</b>	<b>Kat.Nr./Cat.No. : 42533</b>
<b>Sample Buffer for Blue Native (2x)</b>	<b>Lot/Contr.No. : P190360</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>pH</b>	6.9 – 7.1	7.1
<b>Zusammensetzung</b> composition		
ε-Aminocapronsäure/ ε-Aminocaproic acid (#12548)	1 M	entspricht corresponds
Bis(2-hydroxyethylamino]-2-(hydroxy-methyl)- 1,3-propanediol/ BISTRIS (#15107)	100 mM	entspricht corresponds
Natriumchlorid/ Sodium chloride (#30183)	100 mM	entspricht corresponds
Glycerin/ Glycerol (#23176)	20 %	entspricht corresponds
Serva Blau G250/ Serva Blue G250 (#35050)	0.2 %	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		28.10.2021
<b>Lagerung (°C)</b> storage		-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 29.10.2019**

Christian Monsler

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.