

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Probenpuffer für Clear Native (2x)	Kat.Nr./Cat.No. : 42534
Sample Buffer for Clear Native (2x)	Lot/Contr.No. : P200131

Parameter parameter	Spezifikation specification	Ergebnis result
pH	6.9 – 7.1	7.1
Zusammensetzung composition		
ε-Aminocapronsäure/ ε-Aminocaproic acid (#12548)	4 mM	entspricht corresponds
Imidazol/ Imidazole (#26081)	100 mM	entspricht corresponds
Natriumchlorid/ Sodium chloride (#30183)	100 mM	entspricht corresponds
Glycerin/ Glycerol (#23176)	20 %	entspricht corresponds
Ethylenediamine tetraacetic acid/ EDTA (#11278)	2 mM	entspricht corresponds
Ponceau S (#33429)	0.02 %	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		06.04.2023
Lagerung (°C) storage		- 15 bis/ to - 25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.04.2020

Christian Monsler

Marina Konter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.