

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>TAE Puffer (50x)</b>	<b>Kat.Nr./Cat.No. : 42549</b>
TAE Buffer (50x) molecular biology grade	<b>Lot/Contr.No. : P190012</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
<b>pH</b>	8.1 – 8.5	8.5
<b>Zusammensetzung</b> composition		
TRIS (# 37180)	242.3 g/l	entspricht corresponds
EDTA-Na <sub>2</sub> ·2H <sub>2</sub> O (# 11280)	18.6 g/l	entspricht corresponds
Eisessig/ Acetic acid	60.05 g/l	entspricht corresponds
<b>DNase/ RNase Aktivität</b> DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		09.01.2021
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 14.01.2019**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.