

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

SERVA Tris-Tricin/SDS Proben Puffer (2x)	Kat.Nr./Cat.No. : 42551
SERVA Tris-Tricine/SDS sample buffer (2x)	Lot/Contr.No. : P160320

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	dunkelblau violette Lösung dark blue violet solution	entspricht corresponds
pH	8.0 – 9.0	8.4
Zusammensetzung composition		
TRIS/ HCl pH 8.45	900mM	entspricht corresponds
Glycerin/ glycerol (#23176)	24%	entspricht corresponds
SDS (# 20763)	4%	entspricht corresponds
SERVA Blau G/ Blue G (#35050)	0.015%	entspricht corresponds
Phenolrot/ phenol red (#32095)	0.005%	entspricht corresponds
Anwendung application	für SDS-PAGE for SDS-PAGE	
Mindesthaltbarkeit minimum shelf life		04.08.2023
Lagerung (°C) storage		+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.06.2021

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.