

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA Tris-Tricin/SDS Elektrophorese Puffer (10x)	Kat.Nr./Cat.No. :	42552
SERVA Tris-Tricine/SDS electrophoresis buffer (10x)	Lot/Contr.No. :	P120185

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	7.9 – 8.5	8.5
Zusammensetzung composition		
TRIS (# 37180)	1M	entspricht corresponds
Tricin/ Tricine (# 37195)	1M	entspricht corresponds
SDS (# 20763)	1%	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		04/ 2014
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control**Druckdatum/ printing date: 21.02.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.