

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

„Laemmli“ Puffer 10x	Kat.Nr./Cat.No. : 42556
„Laemmli“ Buffer 10x	Lot/Contr.No. : P250219

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	8.4 – 8.9	8.7
Zusammensetzung composition		
TRIS (# 37180)	0.25 M	entspricht corresponds
Glycin / Glycine (# 23390)	1.92 M	entspricht corresponds
SDS (# 20765)	1 %	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		08.07.2027
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 09.07.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.