

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA Tris-Tricin/SDS Elektrophorese Puffer (20x)	Kat.Nr./Cat.No. :	42560
SERVA Tris-Tricine/SDS electrophoresis buffer (20x)	Lot/Contr.No. :	P210096

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare farblose Lösung clear colorless solution	entspricht corresponds
pH	8.3 – 8.7	8.6
Zusammensetzung composition		
TRIS (# 37180)	1.2 M	entspricht corresponds
Tricin/ Tricine (# 37195)	800 mM	entspricht corresponds
SDS (# 20763)	2 %	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		23.03.2024
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 08.04.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.