

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVALight Helios	Kat.Nr./Cat.No.:	42587
CL HRP WB Substrate Kit	Lot/Contr.No.:	131252 – 131253

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
Anwendung application	Chemiluminesenz chemiluminescence	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			05/ 2015
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 14.11.2013

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.