

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>PBS Puffer (10x), steril</b>	<b>Kat.Nr./Cat.No. : 42595</b>
PBS Buffer (10x), sterile	<b>Lot/Contr.No. : P230305</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
<b>pH</b>	1:10 verdünnt 1:10 diluted	7.2 – 7.6	7.4
<b>Sterilttest</b> sterile test		kein Wachstum no growth	entspricht corresponds

**Zusammensetzung**  
composition

NaCl (# 30183)	1.37 M
KCl (# 26868)	27 mM
Na <sub>2</sub> HPO <sub>4</sub> •2H <sub>2</sub> O (# 30200)	100 mM
KH <sub>2</sub> PO <sub>4</sub> (# 26870)	20 mM

**Mindesthaltbarkeit** 08.08.2025  
minimum shelf life

**Lagerung (°C)** +15 bis/ to +30  
storage

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 22.08.2023**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.