

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

PBST Puffer (10x), steril	Kat.Nr./Cat.No. : 42597
PBST Buffer (10x), sterile	Lot/Contr.No. : P210143

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
pH	1:10 verdünnt 1:10 diluted	7.2 – 7.6	7.4
Steriltest sterile test		kein Wachstum no growth	entspricht corresponds

Zusammensetzung
composition

NaCl (# 30183)	1.37 M
KCl (# 26868)	27 mM
Na ₂ HPO ₄ •2H ₂ O (# 30200)	100 mM
KH ₂ PO ₄ (# 26870)	20 mM
Tween® 20 (#37470)	0.5 %

Mindesthaltbarkeit 28.04.2023
 minimum shelf life

Lagerung (°C) +15 bis/ to +30
 storage

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 10.05.2021

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.