

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVALYT[®] PRECOTES[®] CSF Kit	Kat.Nr./Cat.No. :	42800
for IEF of Cerebrospinal Fluid	Lot/Contr.No. :	P120135

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 245x125mm, 300µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 245x125mm, 300µm		
pH-Servalyt	2% in H ₂ O, RT	6.9 ± 0.3	7.0
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.400 – 0.890	0.629
Proteintrennung protein separation	IEF, 300 µm-Gel, 5°C, 3500 Vh	entspricht Standard conform reference	entspricht conform
	getestet mit CSF-Proben und anschließender Silberfärbung tested with CSF samples and silver staining		entspricht corresponds
Haftung des Gels adhesion of the gel	IEF, 300 µm-Gel, 5°C, 3500 Vh	Haftung ja adhesion yes	ja yes
Mindesthaltbarkeit minimum shelf life			09/ 2013
Lagerung (°C) storage			+2 bis/ to +8

<p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 02.04.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.