

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>SERVALYT® PRECOTES® pH 6-9</b>	<b>Kat.Nr./Cat.No. : 42878</b>
<b>300µm, 125x125mm</b>	<b>Lot/Contr.No. : P170378</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 125x125mm, 300µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 125x125mm, 300µm		
<b>pH-Servalyt</b>	2% in H <sub>2</sub> O, RT	7.65 ± 0.4	7.80
<b>Leitfähigkeit ( mS/cm )</b> conductivity	2% in H <sub>2</sub> O, RT	0.200 - 0.650	0.419
<b>Proteintrennung</b> protein separation	IEF, 300 µm-Gel, 5°C, 5000 Vh	entspricht Standard conform reference	entspricht conform
<b>Haftung des Gels</b> adhesion of the gel	IEF, 300 µm-Gel, 5°C, 5000 Vh	Haftung ja adhesion yes	ja yes
<b>Mindesthaltbarkeit</b> minimum shelf life			21.05.2019
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 23.11.2017**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.